

Response To Pre-Bid Queries (Pre-Bid date: 03.10.2022)				
Bid Number: GEM/2022/B/2577897 Dated: 26-09-2022				
Medium end Multipara Monitor				
Sr. No.	Buyer uploaded ATC document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
1	Pg 1 Para 2	Should have integrated medical grade touch screen of at least 15" with display of 10 or more waveforms. Also, should have audio, visual and graded alarming system.	Should have integrated medical grade touch screen of at least 12" or more with display of 10 or more waveforms. Also, should have audio, visual and graded alarming system.	<b>Amended as:</b> Should have integrated medical grade touch screen of at least <b>12" or more</b> with display of 10 or more waveforms. Also, should have audio, visual and graded alarming system.
2	Pg 1 Para 3	Should have 24 hours graphical and numerical trend with split screen facility of all parameters with at least 15 critical alarms summary.	Should have 24 hours graphical and numerical trend with tabular form of all parameters with at least 15 critical alarms summary. Remark/Request/Suggestion: Split screen is not required as the tabular trend table is sufficient to get the complete data for analysis. Should have 24 hours graphical and numerical trend of all parameters with at least 15 critical alarms event	No Change
3	Pg 1 Para 7	Should measure and display Pulse Pressure Variation or Systolic Pressure Variation on main screen.	Kindly remove this point - For wider participation	<b>Deleted</b>
4	Pg 1 Para 10	Monitor should be capability of upgrading to advance modules like BIS and Cardiac Output (Invasive or non-invasive)	Should have drug and Hemodynamics calculation. Remark/Request/Suggestion: Maximum Companies do not have oxygenation and ventilation calculation as this particular thing eliminates maximum participants from the bid. Kindly remove this point - For wider participation	No Change
5	Pg 1 Para 11	Monitor should have capability to interface with devices like Ventilators, anesthesia machine, for display of said device's parameters on Patient Monitor.	Request to DELETE Justification: Company specific parameter available with one or two company, which is limited and restricting the competition, for wider and better participation we request to Delete this. Removed Remark/Request/Suggestion: To make this point into consideration, prototype is required to integrate the connectors of External devices like Ventilators, anesthesia machines etc. Moreover it is a kind of monopoly thing and hence lower the chances of maximum participation of other companies. Kindly remove company specific terms for wider participation. The bid is reserved for Make in India But Point No; 11 is not available with Make in India	<b>Deleted</b>

Response To Pre-Bid Queries (Pre-Bid date: 03.10.2022)				
Bid Number: GEM/2022/B/2577897 Dated: 26-09-2022				
Medium end Multipara Monitor				
Sr. No.	Buyer uploaded ATC document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
6	NA	Suggested by Bidder	Monitor should have trend storage on PEN drive / SD Card. And the data should be easily accessible in computer without help of any special software - For MRD it is very beneficial.	No Change
7	NA	Suggested by Bidder	Should have USB connectivity directly to laser printer - so that if required a plain paper prinout can be taken.	No Change
8	NA	NA	Please clarify regarding Etco2 module whether it is required as a standard parameter of Monitor OR Monitor should have capability of upgrading with Etco2 in future if required by the user.	<b>Clarification:</b> Required as standard parameter
9	NA	NA	Please clarify that the prices of advance modules for upgrading the Monitors like BIS and Cardiac Output (Invasive or non-invasive) should be quoted separately or not.	Not Considered