

Response To Pre-Bid Queries (Pre-Bid date: 30.01.2023)				
Tender Ref:- GEM/2023/B/3031158				
Item Name: High Frequency ventilator				
Sr. No.	Tender Specification Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	TSEC Recommendations
1	Page 1, Para 2	Volume guarantee in every mode	Kindly delete this point. Justification:- Kindly note that the asked specification is available with only few companies like Drager and Vyaire Medical .We request you to kindly delete this point for wider and healthier competition.	No change considered
2	Page 1, 23 (g)	Frequency Up to 200 BPM	Frequency 0 – 150 BPM. Justification:- Kindly note that point 7(c) i.e Frequency :5-150 and the mentioned point 23(g) is contradictory itself. We request you to amend point 23(g) as requested.	Amended as:- Para c. frequency 5 to 200 BPM. Para 23 (g) DELETED