

Response To Pre-Bid Queries (Pre-Bid date: 30.01.2023)				
Tender Ref:- GEM/2022/B/3027660				
Item Name: Defibrillator with CPR monitoring and Transcutaneous pacing				
Sr. No.	Tender Specification Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	TSEC Recommendations
1	Page 1	Defibrillator with CPR monitoring and Transcutaneous pacing	Defibrillator with Transcutaneous pacing Justification:- The asked point is restricting the competition as it is lockout specs available with only few companies. Hence, we request you to kindly amend this point for better and wider participation.	No change considered
2	Page 1, Para 2	Defibrillator Should be biphasic waveform with 3 or more waveform Display.	It should be amend as the defibrillator should be Biphasic wave from with at least 4 or more wave from.	No change considered
3	Page 1 Para 3	Screen size 6" or more diagonal	It should be amend as defibrillator Screen size of Minimum 8" or more Display to all Parameter.	No change considered
4	Page 1, Para 7	It should have ability to measure chest compression rate and depth in real time with both visual & audible feedback. Reusable CPR feedback sensor or similar reusable product - 1 no If disposable, CPR feedback sensor it should be 10 nos	Kindly delete this point. The asked point is restricting the competition as it is lockout specs available with only few companies i.e. Zoll, Stryker and Nihon Kohden. Hence, we request you to kindly delete this point for better and wider participation.	No change considered
5	Page 1, Para 10	It should have ability to filter out CPR artifacts and allowing person to see organized rhythms without interrupting chest compression	Kindly delete this point. Justification:-The asked point is restricting the competition as it is lockout specs available with only few companies i.e. Zoll, Stryker and Nihon Kohden. Hence, we request you to kindly delete this point for better and wider participation.	Amended as:- It should have ability to filter out CPR artifacts to show the actual rhythm.
			It should be deleted its company specific	

6	Page 1, BOQ-1	Defibrillator with CPR monitoring and Transcutaneous pacing- Qty -1	Defibrillator with Transcutaneous pacing- Qty -1 Justification:- The asked point is restricting the competition as it is lockout specs available with only few companies. Hence, we request you to kindly amend this point for better and wider participation.	No change considered
7	Page 2, BOQ-10	Reusable CPR feedback sensor or similar reusable product - 1 no If disposable, CPR feedback sensor - 10 nos	Kindly delete this point. Justification:- The asked accessories are the part of CPR monitoring which is available with just few companies. We request you to kindly delete this point for better and wider participation	No change considered