**07-06-2022**

# Amendment no. 4

**Sub: Technical Specification Amendment for the tender of IIT-Kharagpur.**

**Ref.: GeM Bid No GEM/2022/B/2093051 Dated: 07-04-2022 for item 'OT Table with Orthopedic attachment (Qty: 02 nos.)'**

The following changes are being incorporated in the above referred GeM Bid Number only.

| Tender Para | EXISTING TENDER SPECIFICATION | READ AS |
|---|---|---|
| **5** | Should be able to slide longitudinally more than 250 mm both side | **DELETED** |
| **7** | Two foldable and detachable radiolucent traction bars fixed beneath the seat plate with two adjustable pivot joints. | Two detachable radiolucent traction bars fixed beneath the seat plate with two adjustable pivot joints. |
| **12f.** | Battery capacity for approximately 2 weeks with average use | Battery capacity for approximately **01** weeks with average use |
| **13e.** | Traction bars radiolucent-02 | **Added para:-**<br>**01 trolley for radiolucent traction bars.** |
| **13h.** | Traction boot small pair with multiplanner rotation | **DELETED** |

All other contents of the Bid Document including terms & conditions remain unaltered.

**Note:** Prospective Bidders are also advised to check the GeM Portal regularly prior to the closing date and time of online submission of bid.