07-06-2022

Amendment no. 4

Sub: Technical Specification Amendment for the tender of IIT-Khargapur.

Ref.: GeM Bid No: GEM/2022/B/2092386 Dated: 07-04-2022 for item 'Full Fowler electrical beds with accessories (Qty: 56 nos.)'

The following changes are being incorporated in the above referred GeM Bid Number only.

Tender Para	EXISTING TENDER SPECIFICATION	READ AS
	Full Fowler electrical beds with accessories	**Fully Motorized ICU bed with accessories**
4	Profile frame with easy removable ABS plastic covers	**04 section table top with 4-5" thick multi-layered mattress.**
7	Electrostatic painted metal fraine	Electrostatic painted metal fraine/ **foxy powder coated metal frame**
11	Plastie crash bumpers	Plastic/**rubber crash** bumpers
15	Height range: 47 cm	Height Range:- **470 to 770 mm (without mattress) with tolerance of +/-5%**
22	Should be supplied with other essential accessories	**1. 4-5" thick multi-layered mattress - 1 no** **2. Antimicrobial and flame resistant mattress cover - 1 no** **3. SS 304 grade saline stand - 2 nos** **4. Urinary bag/drain bag holder - 1 no**
23	Manufacturer should be ISO 13485 certified	Manufacturer should have ISO 13485 certification **issued from:** **Ay certification bodies registered with NABCB under Medical Devices quality management system** **OR** **Ay notified body registered with CDSCO** **OR** **Any 4 digit European CE notified body**

All other contents of the Bid Document including terms & conditions remain unaltered.

Note: Prospective Bidders are also advised to check the GeM Portal regularly prior to the closing date and time of online submission of bid.