07-06-2022

Amendment no. 4

Sub: Technical Specification Amendment for the tender of IIT-Khargapur.

Ref.: GeM Bid No GEM/2022/B/2116323 Dated: 19-04-2022 for item 'Platelet Apheresis Machine (Qty: 02 nos.)'

The following changes are being incorporated in the above referred GeM Bid Number only.

EXISTING TENDER SPECIFICATION	READ AS
Added Para	**Blood flow should have range 10ml-120 ml per minute**

All other contents of the Bid Document including terms & conditions remain unaltered.

Note: Prospective Bidders are also advised to check the GeM Portal regularly prior to the closing date and time of online submission of bid.