07-06-2022

Amendment no. 4

Sub: Technical Specification Amendment for the tender of IIT-Kharagpur.

Ref.: GeM Bid No GEM/2022/B/2092614 Dated: 07-04-2022 for item 'ABG Machine (Qty: 01 no.)'

The following changes are being incorporated in the above referred GeM Bid Number only.

Tender Para	EXISTING TENDER SPECIFICATION	READ AS
6	Maintenance free electrodes with individual electrodes ON/OFF facility.	Maintenance free / **Low maintenance** electrodes with individual electrodes ON/OFF facility
10	Data print out on built in graphic printer.	Data print out on built in graphic / **thermal** printer
16	Should be USFDA or European CE with 4 digit certified body no. or BIS approved product	Should be USFDA or European CE IVD or BIS approved product

All other contents of the Bid Document including terms & conditions remain unaltered.

Note: Prospective Bidders are also advised to check the GeM Portal regularly prior to the closing date and time of online submission of bid.