07-06-2022

Amendment no. 4

Sub: Technical Specification Amendment for the tender of IIT-Kharagpur.

Ref.: GeM Bid No GEM/2022/B/2116648 Dated: 19-04-2022 for item 'Laparoscope instrument Set & Laparoscopy hardware (Qty: 02 nos.)'

The following changes are being incorporated in the above referred GeM Bid Number only.

Tender Para	EXISTING TENDER SPECIFICATION	READ AS
3.1	CAMERA CONTROL UNIT & CAMERA	
a	3 x 1/3 Progressive scan CCD or 3 x 1/3 Progressive scan CMOS	3 chip full HD camera (CCD or CMOS) with progressive scan
3.4	CO2 Insufflator	
2	Adjustable flow rate of 45 litres per minute or more and pressure range adjustable between 0 to 30 mm Hg	Adjustable flow rate of 45 litres per minute or more and pressure range adjustable between **3 to 25 mm Hg**

All other contents of the Bid Document including terms & conditions remain unaltered.

Note: Prospective Bidders are also advised to check the GeM Portal regularly prior to the closing date and time of online submission of bid.