

Amendment No. 02

Date: 30.01.2024

Sub: Amendment No.02 to the Tender Enquiry Document

Ref: (i) Tender No: HITES/PCD/AIIMS/PET-CT/23-24 Dated 03-01-2024

Section VII
Technical Specifications

TECHNICAL AMENDMENT- HITES/PCD/AIIMS/PET-CT/23-24		
Item Name: PET CT		
Tender Page & Para	TENDER SPECIFICATION	AMMENDED AS
	Site Modification/ Turnkey Clarification	
Page no. 51, Site modification, Point no II	Supplier shall ensure that the equipment model quoted is commissioned within the designated area with the area constructed by the Hospital as per the provided AERB approved layout plan. All minor Structural changes which are part of turnkey including any brick work or concrete work should be carried as per AERB approved drawing	Supplier shall ensure that the equipment model quoted is commissioned within the designated area with the area constructed by the Hospital as per the provided AERB approved layout plan. All minor Structural changes which are part of turnkey including any brick work should be carried as per AERB approved drawing. Bidder shall be also responsible for construction of concrete walls if required. The density and thickness should be as per AERB guidelines. The rate for construction of concrete walls as per AERB guidelines for 100 cubic metre should be quoted separately and shall be considered for price comparison. The payment shall be made as per actual measurement basis.

All other terms and conditions of the tender enquiry remain unaltered