

Amendment No. 7

25.01.2023

Sub: Amendment to the Global Tender Enquiry Document**Ref.: Notice Inviting Bid ref. HITES/PCD/MCH-AIIMS/02/22-23 dated 24.09.2022 read with its Amendment no. 01, 02, 03, 04, 05 & 06 dated 20.10.2022, 17.11.2022, 05.12.2022, 21.12.2022, 03.01.2022 & 10.01.2023 respectively and Corrigendum 2 for Auto Bid Extension.**

The following changes have been authorised and are being incorporated in the above referred Bidding Document for the below mentioned tender only:

- i)
- CT 256 Slice (Dual Energy) (Tender ID: 2022_HLL_129553_1),**

SECTION – VII**TECHNICAL SPECIFICATION AND GENERAL POINTS****A. TECHNICAL SPECIFICATION:****Item No. 1 (Tender ID: 2022_HLL_129553_1)****CT 256 Slice (Dual Energy)**

Sr. No.	Tender document Para & Page	Exiting Tender Specification	Amended As
1	Uptime & Downtime penalty: Pg. 89 & 90.	During the Warranty period and CAMC period, desired Uptime of 95% of 365/366 (Leap Year) days (24 hrs), if downtime more than 5%, the warranty period/CAMC period will be extended by double the downtime period. In addition a penalty equal to amount of 0.25 % of the total cost of equipment per day will be liveable for the excess downtime period. Complaints should be attended properly, maximum within 8 hrs.	During the Warranty period and CAMC period, desired Uptime of 95% of 365/366 (Leap Year) days (24 hrs), if downtime more than 5%, the warranty period/CAMC period will be extended by double the downtime period. Complaints should be attended properly, maximum within 8 hrs.

Sr. No.	Tender document Para & Page	Exiting Tender Specification	Amended As
2	List of Requirement Part II: Required Delivery Schedule: Point a & b. Pg. 43 of TED.	Installation and Commissioning shall be done at the earliest but not later than 15 days of delivery of goods at site or date of handing over the site for installation, whichever is later.	Installation and Commissioning shall be done at the earliest but not later than 45 (forty Five) days of delivery of goods at site or date of handing over the site for installation, whichever is later.

All other contents of the Global Tender Enquiry Document including terms & conditions remain unaltered.