

**Amendment No 16**

Date: 03.04.2024

Sub: Amendment No.16 to the Tender Enquiry Document

Ref: (i) Tender No: HITES/PCD/AIIMS JODHPUR/CYCLOTRON/RT-02/23-24 Dated 26-10-2023

TENDER ID- 2023\_HITE\_172179\_1

<b>Technical Amendment- Tender for procurement of Cyclotron for AIIMS, Jodhpur</b>		
<b>Tender Ref. HITES/PCD/AIIMS JODHPUR/CYCLOTRON/RT 02/23 24 Dated:26.10.2023</b>		
<b>Item Name : Medical Cyclotron</b>		
<b>(Clause No &amp; Page No)</b>	<b>Tender Specification</b>	<b>Amended As</b>
Page No 46 point 1 General, sub point VI	VI. Cyclotron should be CE certified and should be type approved in the country of origin and The company shall comply and document with all the regulatory requirements of AERB	Cyclotron and all accessories in the Medical Cyclotron Facility setup should be CE certified and also certified by the Nuclear Safety Commission from the country of origin of the OEM. Cyclotron should be type approved in the country of origin and the company shall comply and document with all the regulatory requirements of AERB.

**All other terms and conditions of the tender enquiry remain unaltered**