Response To Pre-Bid Queries (Pre-Bid date: 13.03.2023)				
Bid Number: GEM/2023/B/3223813 Dated: 04-03-2023				
Intermediate anaesthesia machine with multipara monitor				
Sr. No.	**Buyer Specification Document Page & Para**	**TENDER SPECIFICATION**	**REPRESENTATION RECEIVED FROM THE FIRM**	**COMMITTEE RECOMMENDATION**
1	Page 1 Para 4 a)	Should have fresh gas de-coupled semi closed circle absorber system.	Should have fresh gas de-coupled/fresh gas compensation semi closed circle absorber system	**Amended as:** Should have fresh gas de-coupled **/ fresh gas compensation** semi closed circle absorber system
2	Page 1 Para 5 g)	PEEP: 0 ~ 20 mbar	PEEP: off, 3 ~ 20 mbar	No Change
3	Page 1 Para 5 l)	Should be able to ventilate with atmospheric air, in case of total gas supply failure.	It should be deleted as it is company specific.	**Deleted**
4	Page 1 Para 9	Specification of Anaesthesia Gas Module (price to be quoted as optional)	Please confirm that AGM required as optional or as standard.	**Amended as:** Specification of Anaesthesia Gas Module
5	Page 1 Para 9 b)	Machine should also have the facility to monitor side stream ETCO2 and should display the same on the patient monitor attached to the anesthesia machine.	It should be deleted as it is company specific.	No Change
6	Page 2 Para 9 d)	Anesthesia Gas Monitor should be from the same OEM from which the anesthesia machine is procured.	Anesthesia Gas Monitor should be from the same OEM or OEM integrated from which the anesthesia machine is procured.	No Change
7	Page 2 Para 11 2	Should have integrated medical grade touch screen of at least 15" with display of 10 or more waveforms. Also, should have audio, visual and graded alarming system.	Should have integrated medical grade touch screen of at least 12" with display of 10 or more waveforms. Also, should have audio, visual and graded alarming system.	No Change
8	Page 2 Para 11 10	Monitor should be capability of upgrading to advance modules like BIS and Cardiac Output (Invasive or non-invasive)	It should be deleted because as these items are available from other companies and will not support MII clause as required in the bid.	No Change