Response To Pre-Bid Queries (Pre-Bid date: 13.03.2023)				
Bid Number: GEM/2023/B/3223447 Dated: 04-03-2023				
ECMO				
Sr. No.	**Buyer Specification Document Page & Para**	**TENDER SPECIFICATION**	**REPRESENTATION RECEIVED FROM THE FIRM**	**COMMITTEE RECOMMENDATION**
1	Page 1 Para A. b.	Should be able to flow 0 to 9.99 LPM.	Flow should be 0.8 Lits/Min and above	**Amended as:** Flow should be **0.8 Lits/Min and above**
2	Page 1 Para A. f. 5	01 no. Venous Blood Oxygen Saturation (SvO2) (measuring Range – 40 to 100%).	To be removed since it is not part of ECMO Equipment	No Change
3	Page 1 Para A. f. 6	01 no. Hemoglobin (Measuring Range 05 to 15 gm/dl).	To be removed since it is not part of ECMO Equipment	No Change
4	Page 1 Para A. f. 7	01 no. Hematocrit (Measuring range 15 to 50%)	To be removed since it is not part of ECMO Equipment	No Change
5	Page 1 Para E. i.	Should have integrated Measuring Cell for Venous probe to measure Venous Oxygen saturation (SvO2), Hemoglobin, Hematocrit and Venous Temperture.	It is repetition of above mentioned point (f) - 5, 6, 7 - so it need to be removed.	No Change