Response To Pre-Bid Queries (Pre-Bid date: 13.03.2023)				
Bid Number: GEM/2023/B/3223203 Dated: 04-03-2023				
Transport basic vital monitor				
Sr. No.	**Buyer Specification Document Page & Para**	**TENDER SPECIFICATION**	**REPRESENTATION RECEIVED FROM THE FIRM**	**COMMITTEE RECOMMENDATION**
1	Page 1 Para 1	High – resolution colour display of upto 10 inches.	High – resolution colour display of upto **12.1 inches.** Justification: Bigger the screen size better the visibility will be. Kindly amend this point for wider participitation	No Change
2	Page 1 Para 13	Battery backup for at least 4 Hrs.	Battery back up for at least 3.5hrs or more.	**Amended as:** Battery back up for at least **3.5 Hrs or more.**
3	NA	Suggested by Bidder	Monitor should have trend storage of 168 hours Justification: This will help in storing more information.	No Change
4	NA	Suggested by Bidder	Monitor should have trend storage on PEN drive / SD Card. And the data should be easily accessible in computer without help of any special software Justification: for MRD records it would be helpful	No Change
5	NA	Suggested by Bidder	Should have USB connectivity directly to laser printer Justification: so that if required a plain paper printout can be taken.	No Change
6	NA	Suggested by Bidder	Should have 12 lead ECG monitoring facility by 10 lead ECG cable. Justification: 12 lead ECG monitoring will help in proper analysis of ischemia related disease	No Change
7	NA	Suggested by Bidder	Should have minimum 12 waveform on display Justification: Helps in veiwing of all 12 lead ECG in one single screen	No Change
8	NA	Suggested by Bidder	Should have softwares lik ST analaysis, ST mapping, QT analysis, HRV analysis and should measure PVI & PI Justification: for proper diagonosis these features are required in a multiparamonitor	No Change