Response To Pre-Bid Queries (Pre-Bid date: 13.03.2023)				
Bid Number: GEM/2023/B/3223727 Dated: 04-03-2023				
Recovery Monitor				
Sr. No.	**Buyer Specification Document Page & Para**	**TENDER SPECIFICATION**	**REPRESENTATION RECEIVED FROM THE FIRM**	**COMMITTEE RECOMMENDATION**
1	Page 1 Para 1	Modular Unit with high resolution (640 x 480 pixels) colour TFT display at least 15.0" & touch screen.	Modular unit with high resolution (800 x 600 pixels) color TFT display at least 12 inch or more & touch screen. Modular Unit with high resolution (800×600 pixels) colour TFT display at least 12.0" or more & touch screen. Justification: Kindly amend this point for wider participitation	No Change
2	Page 1 Para 4	Monitor should have lithium-ion / nickel cadmium battery back-up facility for at least 1 hrs.	Monitor should have lithium-ion/nickel cadmium battery back-up facility for at least 3.5 hrs.	No Change
3	NA	Suggested by Bidder	Monitor should have trend storage of 168 hours Justification: This will help in storing more information.	No Change
4	NA	Suggested by Bidder	Monitor should have trend storage on PEN drive / SD Card. And the data should be easily accessible in computer without help of any special software Justification: for MRD records it would be helpful	No Change
5	NA	Suggested by Bidder	Should have USB connectivity directly to laser printer Justification: so that if required a plain paper printout can be taken.	No Change
6	NA	Suggested by Bidder	Should have 12 lead ECG monitoring facility by 10 lead ECG cable. Justification: 12 lead ECG monitoring will help in proper analysis of ischemia related disease	No Change
7	NA	Suggested by Bidder	Should have minimum 12 waveform on display Justification: so that if required a plain paper printout can be taken.	No Change
8	NA	Suggested by Bidder	Should have softwares lik ST analaysis, ST mapping, QT analysis, HRV analysis and should measure PVI & PI Justification: for proper diagonosis these features are required in a multiparamonitor	No Change